

表 398 ETRF装着率(%) 細菌数測定の種類別（ベッドサイドコンソール台数≧1）

細菌数測定の種類	0(装着せず)	10未満	10～	20～	30～	40～	50～	60～	70～	80～	90～	100(全台装着)	合計	記載なし	総計	平均	標準偏差
なし	57	7	8	5	3	2	6	5	3	2	4	139	241	1	242	65.50	44.32
(%)	(23.7)	(2.9)	(3.3)	(2.1)	(1.2)	(0.8)	(2.5)	(2.1)	(1.2)	(0.8)	(1.7)	(57.7)	(100.0)				
毎日	1					1				1		13	16		16	89.15	28.17
(%)	(6.3)					(6.3)				(6.3)		(81.3)	(100.0)				
毎週		2		2		2	3	2	3	4	9	100	127	1	128	93.63	18.00
(%)		(1.6)		(1.6)		(1.6)	(2.4)	(1.6)	(2.4)	(3.1)	(7.1)	(78.7)	(100.0)				
隔週	3	4	1	3	4	3	2	1	6	9	13	211	260		260	92.75	21.05
(%)	(1.2)	(1.5)	(0.4)	(1.2)	(1.5)	(1.2)	(0.8)	(0.4)	(2.3)	(3.5)	(5.0)	(81.2)	(100.0)				
毎月	54	48	47	44	42	25	38	36	42	56	119	2,192	2,743	1	2,744	90.17	25.15
(%)	(2.0)	(1.7)	(1.7)	(1.6)	(1.5)	(0.9)	(1.4)	(1.3)	(1.5)	(2.0)	(4.3)	(79.9)	(100.0)				
年数回	17	6	5	4	6	2	7	10	6	12	23	324	422		422	88.88	26.82
(%)	(4.0)	(1.4)	(1.2)	(0.9)	(1.4)	(0.5)	(1.7)	(2.4)	(1.4)	(2.8)	(5.5)	(76.8)	(100.0)				
年1回	14	3	9	7	6	1	4	5	7	12	16	309	393	2	395	89.26	26.73
(%)	(3.6)	(0.8)	(2.3)	(1.8)	(1.5)	(0.3)	(1.0)	(1.3)	(1.8)	(3.1)	(4.1)	(78.6)	(100.0)				
合計	146	70	70	65	61	36	60	59	67	96	184	3,288	4,202	5	4,207	88.80	27.16
(%)	(3.5)	(1.7)	(1.7)	(1.5)	(1.5)	(0.9)	(1.4)	(1.4)	(1.6)	(2.3)	(4.4)	(78.2)	(100.0)				
不明	13	3	2	6	1	1	3	3	2	1	4	49	88	2	90	69.36	41.00
(%)	(14.8)	(3.4)	(2.3)	(6.8)	(1.1)	(1.1)	(3.4)	(3.4)	(2.3)	(1.1)	(4.5)	(55.7)	(100.0)				
記載なし												2	2	5	7	100.00	
(%)												(100.0)	(100.0)				
総計	159	73	72	71	62	37	63	62	69	97	188	3,339	4,292	12	4,304	88.41	27.64
(%)	(3.7)	(1.7)	(1.7)	(1.7)	(1.4)	(0.9)	(1.5)	(1.4)	(1.6)	(2.3)	(4.4)	(77.8)	(100.0)				

数値下のかっこ内は行方向の合計に対する%です。

(C)Japanese Society for Dialysis Therapy

2014年調査