

表 403 ETRF装着率(%) ET測定頻度別（ベッドサイドコンソール台数≧1）

ET測定頻度	0(装着せず)	10未満	10～	20～	30～	40～	50～	60～	70～	80～	90～	100(全台装着)	合計	記載なし	総計	平均	標準偏差
なし	33	1	2	4	1	2	3	1	7		1	62	117		117	62.18	44.94
(%)	(28.2)	(0.9)	(1.7)	(3.4)	(0.9)	(1.7)	(2.6)	(0.9)	(6.0)		(0.9)	(53.0)	(100.0)				
毎日										1	2	18	21		21	98.97	3.26
(%)										(4.8)	(9.5)	(85.7)	(100.0)				
毎週	1	1		1	1	1	3	1	2	3	5	140	159		159	95.91	15.37
(%)	(0.6)	(0.6)		(0.6)	(0.6)	(0.6)	(1.9)	(0.6)	(1.3)	(1.9)	(3.1)	(88.1)	(100.0)				
隔週	2	1		2	2	2	5	2	4	3	16	229	268		268	95.52	15.99
(%)	(0.7)	(0.4)		(0.7)	(0.7)	(0.7)	(1.9)	(0.7)	(1.5)	(1.1)	(6.0)	(85.4)	(100.0)				
毎月	48	43	37	38	23	40	35	40	45	60	139	2,427	2,975	1	2,976	91.79	22.81
(%)	(1.6)	(1.4)	(1.2)	(1.3)	(0.8)	(1.3)	(1.2)	(1.3)	(1.5)	(2.0)	(4.7)	(81.6)	(100.0)				
年数回	11	9	7	5	3	4	5	5	9	10	15	284	367		367	88.38	27.33
(%)	(3.0)	(2.5)	(1.9)	(1.4)	(0.8)	(1.1)	(1.4)	(1.4)	(2.5)	(2.7)	(4.1)	(77.4)	(100.0)				
年1回	14	4	5	6	3	5	3	3	7	9	12	254	325		325	88.58	27.61
(%)	(4.3)	(1.2)	(1.5)	(1.8)	(0.9)	(1.5)	(0.9)	(0.9)	(2.2)	(2.8)	(3.7)	(78.2)	(100.0)				
合計	109	59	51	56	33	54	54	52	74	86	190	3,414	4,232	1	4,233	90.85	24.41
(%)	(2.6)	(1.4)	(1.2)	(1.3)	(0.8)	(1.3)	(1.3)	(1.2)	(1.7)	(2.0)	(4.5)	(80.7)	(100.0)				
不明	13		2	2	1		1	4		3	1	33	60	3	63	67.67	42.59
(%)	(21.7)		(3.3)	(3.3)	(1.7)		(1.7)	(6.7)		(5.0)	(1.7)	(55.0)	(100.0)				
記載なし					1		1						2	5	7	41.67	11.79
(%)					(50.0)		(50.0)						(100.0)				
総計	122	59	53	58	35	54	56	56	74	89	191	3,447	4,294	9	4,303	90.51	24.92
(%)	(2.8)	(1.4)	(1.2)	(1.4)	(0.8)	(1.3)	(1.3)	(1.3)	(1.7)	(2.1)	(4.4)	(80.3)	(100.0)				

数値下のかっこ内は行方向の合計に対する%です。
(C)Japanese Society for Dialysis Therapy
2015年調査