

表 406 ETRF装着率(%) 細菌数(cfu/mL)別（ベッドサイドコンソール台数≧1）

細菌数	0(装着せず)	10未満	10～	20～	30～	40～	50～	60～	70～	80～	90～	100(全台装着)	合計	記載なし	総計	平均	標準偏差
0.1未満	26	27	15	28	17	22	31	26	44	58	132	2,478	2,904	1	2,905	94.53	18.27
(%)	(0.9)	(0.9)	(0.5)	(1.0)	(0.6)	(0.8)	(1.1)	(0.9)	(1.5)	(2.0)	(4.5)	(85.3)	(100.0)				
0.1～	10	8	11	2	3	8	8	10	10	9	29	380	488	1	489	90.29	24.63
(%)	(2.0)	(1.6)	(2.3)	(0.4)	(0.6)	(1.6)	(1.6)	(2.0)	(2.0)	(1.8)	(5.9)	(77.9)	(100.0)				
1～	13	14	10	11	5	8	4	6	6	11	13	286	387		387	84.81	31.11
(%)	(3.4)	(3.6)	(2.6)	(2.8)	(1.3)	(2.1)	(1.0)	(1.6)	(1.6)	(2.8)	(3.4)	(73.9)	(100.0)				
10～	12	6	8	6	4	7	5	8	3	3	3	96	161		161	73.83	37.01
(%)	(7.5)	(3.7)	(5.0)	(3.7)	(2.5)	(4.3)	(3.1)	(5.0)	(1.9)	(1.9)	(1.9)	(59.6)	(100.0)				
100以上	1				1		2		1	2	1	16	24		24	87.28	25.51
(%)	(4.2)				(4.2)		(8.3)		(4.2)	(8.3)	(4.2)	(66.7)	(100.0)				
合計	62	55	44	47	30	45	50	50	64	83	178	3,256	3,964	2	3,966	92.17	22.27
(%)	(1.6)	(1.4)	(1.1)	(1.2)	(0.8)	(1.1)	(1.3)	(1.3)	(1.6)	(2.1)	(4.5)	(82.1)	(100.0)				
不明	38	3	5	10	2	7	3	6	5	4	11	131	225	2	227	71.50	40.49
(%)	(16.9)	(1.3)	(2.2)	(4.4)	(0.9)	(3.1)	(1.3)	(2.7)	(2.2)	(1.8)	(4.9)	(58.2)	(100.0)				
記載なし	22	1	4	1	3	2	3		5	2	2	60	105	5	110	68.25	42.55
(%)	(21.0)	(1.0)	(3.8)	(1.0)	(2.9)	(1.9)	(2.9)		(4.8)	(1.9)	(1.9)	(57.1)	(100.0)				
総計	122	59	53	58	35	54	56	56	74	89	191	3,447	4,294	9	4,303	90.51	24.92
(%)	(2.8)	(1.4)	(1.2)	(1.4)	(0.8)	(1.3)	(1.3)	(1.3)	(1.7)	(2.1)	(4.4)	(80.3)	(100.0)				

数値下のかっこ内は行方向の合計に対する%です。
(C)Japanese Society for Dialysis Therapy
2015年調査