

表 425 ET測定値(EU/mL) 施設所属患者数別（ベッドサイドコンソール台数≥1）

施設所属患者数	0.001EU/mL未満	0.001～0.01EU/mL未満	0.01～0.05EU/mL未満	0.05～0.1EU/mL未満	0.1～0.25EU/mL未満	0.25～0.5EU/mL未満	0.5EU/mL以上	合計	不明	記載なし	総計
10未満	140	25	14	5	5	2	2	193	33	21	247
(%)	(72.5)	(13.0)	(7.3)	(2.6)	(2.6)	(1.0)	(1.0)	(100.0)			
10～	1,019	188	91	15	12	7	3	1,335	78	32	1,445
(%)	(76.3)	(14.1)	(6.8)	(1.1)	(0.9)	(0.5)	(0.2)	(100.0)			
50～	1,217	180	65	26	10	7	6	1,511	14	8	1,533
(%)	(80.5)	(11.9)	(4.3)	(1.7)	(0.7)	(0.5)	(0.4)	(100.0)			
100～	755	97	33	8	4	2	4	903	4	2	909
(%)	(83.6)	(10.7)	(3.7)	(0.9)	(0.4)	(0.2)	(0.4)	(100.0)			
200～	97	14	3	1	2	1		118	2		120
(%)	(82.2)	(11.9)	(2.5)	(0.8)	(1.7)	(0.8)		(100.0)			
300～	40	5	3				1	49			49
(%)	(81.6)	(10.2)	(6.1)				(2.0)	(100.0)			
合計	3,268	509	209	55	33	19	16	4,109	131	63	4,303
(%)	(79.5)	(12.4)	(5.1)	(1.3)	(0.8)	(0.5)	(0.4)	(100.0)			
記載なし											
(%)											
総計	3,268	509	209	55	33	19	16	4,109	131	63	4,303
(%)	(79.5)	(12.4)	(5.1)	(1.3)	(0.8)	(0.5)	(0.4)	(100.0)			
平均	79.66	72.03	67.12	65.38	60.39	64.26	85.19	77.68	29.76	26.24	75.47
標準偏差	62.27	67.41	72.82	48.31	54.20	57.04	79.56	63.43	35.34	31.02	63.22

数値下のかっこ内は行方向の合計に対する%です。
(C)Japanese Society for Dialysis Therapy
2015年調査