

表 468 専従専門医の有無 ETRF装着率(%) (ベッドサイドコンソール台数≥1)

ETRF装着率	なし	あり	合計	記載なし	総計
0(装着せず)	77	45	122		122
(%)	(63.1)	(36.9)	(100.0)		
10未満	27	32	59		59
(%)	(45.8)	(54.2)	(100.0)		
10~	23	30	53		53
(%)	(43.4)	(56.6)	(100.0)		
20~	35	23	58		58
(%)	(60.3)	(39.7)	(100.0)		
30~	10	25	35		35
(%)	(28.6)	(71.4)	(100.0)		
40~	32	22	54		54
(%)	(59.3)	(40.7)	(100.0)		
50~	24	32	56		56
(%)	(42.9)	(57.1)	(100.0)		
60~	33	22	55	1	56
(%)	(60.0)	(40.0)	(100.0)		
70~	41	31	72	2	74
(%)	(56.9)	(43.1)	(100.0)		
80~	39	50	89		89
(%)	(43.8)	(56.2)	(100.0)		
90~	92	99	191		191
(%)	(48.2)	(51.8)	(100.0)		
100(全台装着)	1,736	1,701	3,437	10	3,447
(%)	(50.5)	(49.5)	(100.0)		
合計	2,169	2,112	4,281	13	4,294
(%)	(50.7)	(49.3)	(100.0)		
記載なし	5	3	8	1	9
(%)	(62.5)	(37.5)	(100.0)		
総計	2,174	2,115	4,289	14	4,303
(%)	(50.7)	(49.3)	(100.0)		
平均	89.96	91.05	90.50	93.09	90.51
標準偏差	25.80	24.03	24.94	13.40	24.92

数値下のかっこ内は行方向の合計に対する%です。

(C)Japanese Society for Dialysis Therapy

2015年調査