

表 488 細菌数測定の頻度 細菌数測定サンプル量別 (ベッドサイドコンソール台数 1)

細菌数測定サンプル量	なし	毎日	毎週	隔週	毎月	年数回	年1回	合計	不明	記載なし	総計
1mL未満			8	12	199	47	23	289	1		290
(%)			(2.8)	(4.2)	(68.9)	(16.3)	(8.0)	(100.0)			
1～10mL未満		2	5	17	337	67	55	483	1		484
(%)		(0.4)	(1.0)	(3.5)	(69.8)	(13.9)	(11.4)	(100.0)			
10～50mL未満		3	35	43	874	106	126	1,187	1		1,188
(%)		(0.3)	(2.9)	(3.6)	(73.6)	(8.9)	(10.6)	(100.0)			
50～100mL未満		3	71	121	1,111	90	161	1,557			1,557
(%)		(0.2)	(4.6)	(7.8)	(71.4)	(5.8)	(10.3)	(100.0)			
100～500mL未満		5	28	34	333	37	52	489			489
(%)		(1.0)	(5.7)	(7.0)	(68.1)	(7.6)	(10.6)	(100.0)			
500mL～1L未満			1	1	14	1	2	19			19
(%)			(5.3)	(5.3)	(73.7)	(5.3)	(10.5)	(100.0)			
1～10L未満				1	8		2	11			11
(%)				(9.1)	(72.7)		(18.2)	(100.0)			
10L以上					3		1	4			4
(%)					(75.0)		(25.0)	(100.0)			
合計		13	148	229	2,879	348	422	4,039	3		4,042
(%)		(0.3)	(3.7)	(5.7)	(71.3)	(8.6)	(10.4)	(100.0)			
不明	114				7	6	3	130	75		205
(%)	(87.7)				(5.4)	(4.6)	(2.3)	(100.0)			
記載なし	70							70		1	71
(%)	(100.0)							(100.0)			
総計	184	13	148	229	2,886	354	425	4,239	78	1	4,318
(%)	(4.3)	(0.3)	(3.5)	(5.4)	(68.1)	(8.4)	(10.0)	(100.0)			

数値下のかっこ内は行方向の合計に対する%です。

##C0065##

(C)Japanese Society for Dialysis Therapy

2016年調査